



**MANTECH  
ELECTRONICS**

*International Components Distributor*

Vat Reg.: 4850196108 – Co. Reg. 2001/004235/07

Address: 32 Laub Street  
New Centre  
Johannesburg  
Phone: 011 493-9307  
Fax: 011 493-9319  
Web: [www.mantech.co.za](http://www.mantech.co.za)  
E-mail: [jn@mantech.co.za](mailto:jn@mantech.co.za)



**ALCOSCAN AL-7000**

### Specifications

Category	Contents
Indication of B.A.C	0.00 to 0.40% B.A.C. or mg/l equivalent by LED display
Accuracy	+/-0.010% BAC at 0.050% BAC
Mouth piece	5 supplied with unit and additional mouthpieces available
Sensor	Semiconductor type alcohol sensor
Power supply	Two 1.5V "AA" Alkaline batteries
Weight	103g (Including batteries)
Dimension (mm)	116 (height) X 50 (width) X 21
Remarks	<ul style="list-style-type: none"><li>- Auto power off</li><li>- Battery Low indication</li><li>- CE, DOT, FDA approval</li><li>- Air flow rate checking</li><li>- Calibrated to legal limit of 0.05 % B.A.C</li><li>- Self sensor testing</li><li>- Easy to replace sensor</li><li>- Features the ability to change sensors on one's own.</li></ul> <p><b>WE RECOMMEND THAT YOU ADD A PACK OF 100 MOUTHPIECES</b></p>